

IFN- γ GoStix™ Plus

Catalog Nos.	Amount	Lot Number
631285	20 tests	Specified on product label.
631286	50 tests	Specified on product label.

Description

IFN- γ GoStix Plus are designed to quantify the levels of human IFN- γ in cell culture supernatant or serum samples in just 10 minutes from 20 μ l of input material, providing accuracy and consistency comparable to an ELISA in a fraction of the time. Following sample application to the GoStix cassette and a brief incubation period, quantification is performed by scanning test and control bands on the cassette using the GoStix Plus smartphone app, which references a lot-specific standard curve to deliver assay results in pg/ml. The included IFN- γ Control allows for confirmation of IFN- γ GoStix Plus functionality. Visit takarabio.com/gostixhelp for more information about GoStix assays and additional guidance on how to download and use the smartphone app.

Package Contents

631285 (20 tests):

- 20 x IFN- γ GoStix Plus (1 test each)
- 2 x IFN- γ Chase Buffer (3 ml/vial)
- 1 x IFN- γ Control (5 tests/vial)

631286 (50 tests):

- 50 x IFN- γ GoStix Plus (1 test each)
- 4 x IFN- γ Chase Buffer (3 ml/vial)
- 1 x IFN- γ Control (5 tests/vial)

Storage Conditions

- Store all components at room temperature upon receipt.
- After reconstitution of the IFN- γ Control, it can be stored at 4°C for up to 1 week. For extended storage, prepare aliquots to store at -20°C and avoid repeated freeze-thaw cycles.

Shelf Life

- 1 year from date of receipt under proper storage conditions.

Shipping Conditions

- Room temperature

Assay Range

- 15 ng/ml to 375 ng/ml

Takara Bio USA, Inc.

2560 Orchard Parkway, San Jose, CA 95131, USA

U.S. Technical Support: technical_support@takarabio.com

United States/Canada
800.662.2566
(012022)

Asia Pacific
+1.650.919.7300

Europe
+33.(0)1.3904.6880

Japan
+81.(0)77.565.6999

Calibration

This assay was calibrated against human recombinant IFN- γ protein. Results obtained with this assay (pg/ml) can be converted to relative approximate NIBSC units (IU/ml). For reference standard 82/587, multiply assay results by 4.96×10^{-4} .

Product Documents

Documents for our products are available for download at takarabio.com/manuals

The following documents apply to this product:

- Cytokine GoStix Plus Protocol-at-a-Glance

Quality Control Data

IFN- γ GoStix Plus were tested in triplicate with a dilution series of recombinant IFN- γ control protein, and samples were read using the GoStix Plus app on both iOS and Android platforms to establish the linearity of the assay (>0.97).

It is certified that this product meets the above specifications, as reviewed and approved by the Quality Department.

IFN-gamma GoStix™ Plus

CATALOG NOS.

631285, 631286

NOTICE TO PURCHASER:

Our products are to be used for **Research Use Only**. They may not be used for any other purpose, including, but not limited to, use in humans, therapeutic or diagnostic use, or commercial use of any kind. Our products may not be transferred to third parties, resold, modified for resale, or used to manufacture commercial products or to provide a service to third parties without our prior written approval.

Your use of this product is also subject to compliance with the licensing requirements, listed below if applicable, and described on the product's web page at <http://www.takarabio.com>. It is your responsibility to review, understand and adhere to any restrictions imposed by these statements.

TRADEMARKS:

©2020 Takara Bio Inc. All Rights Reserved.

All trademarks are the property of Takara Bio Inc. or its affiliate(s) in the U.S. and/or other countries or their respective owners. Certain trademarks may not be registered in all jurisdictions.

Takara Bio USA, Inc.

2560 Orchard Parkway, San Jose, CA 95131, USA

U.S. Technical Support: technical_support@takarabio.com

United States/Canada

Asia Pacific

Europe

Japan

1/20/2022

800.662.2566

+1.650.919.7300

+33.(0)1.3904.6880

+81.(0)77.565.6999