

SMART-Seq® HT Kit

Catalog Nos.	Amount	Lot Number
634455	12 rxns	Specified on product label.
634456	48 rxns	Specified on product label.
634437	96 rxns	Specified on product label.
634438	192 rxns	Specified on product label.

Description

The SMART-Seq HT Kit is designed for the synthesis of high-quality cDNA directly from 1–100 intact cells or ultra-low amounts of total RNA (10–1000 pg). By integrating locked nucleic acid (LNA) and SMART® technologies, this kit offers remarkable sensitivity, enriches for full-length transcripts, and maintains a true representation of the original mRNA pool, which are critical factors for transcriptome analysis. The simplified workflow of this kit makes processing samples easy. cDNA libraries generated with this kit have been tested for compatibility with Illumina® sequencing platforms.

Package Contents

Box 1:

<u>634455</u> (12 rxns)	<u>634456</u> (48 rxns)	<u>634437</u> (96 rxns)	<u>634438</u> (192 rxns)	
5 µl	5 µl	5 µl	2 x 5 µl	Control Total RNA (1 µg/µl)

Box 2:

<u>634455</u> (12 rxns)	<u>634456</u> (48 rxns)	<u>634437</u> (96 rxns)	<u>634438</u> (192 rxns)	
12 µl	48 µl	96 µl	2 x 96 µl	SMART-Seq HT Oligonucleotide
24 µl	96 µl	192 µl	2 x 192 µl	3' SMART-Seq CDS Primer II A
30 µl	75 µl	150 µl	300 µl	RNase Inhibitor (40 U/µl)
125 µl	500 µl	1 ml	2 x 1 ml	One-Step Buffer
24 µl	96 µl	192 µl	2 x 192 µl	SMARTScribe™ Reverse Transcriptase (100 U/µl)
20 µl	20 µl	50 µl	2 x 50 µl	SeqAmp™ DNA Polymerase
230 µl	920 µl	1.85 ml	2 x 1.85 ml	10X Lysis Buffer
1.7 ml	2 x 1.7 ml	6.8 ml	2 x 6.8 ml	Elution Buffer (10 mM Tris-Cl, pH 8.5)
1 ml	2 x 1 ml	4 ml	2 x 4 ml	Nuclease-Free Water

Storage Conditions

- Store Control Total RNA at –70°C
- Store all other reagents at –20°C

Shelf Life

- 1 year from date of receipt under proper storage conditions.

Shipping Conditions

- Dry ice (–70°C)

Takara Bio USA, Inc.

1290 Terra Bella Avenue, Mountain View, CA 94043, USA
 U.S. Technical Support: technical_support@takarabio.com

United States/Canada 800.662.2566 (022421)	Asia Pacific +1.650.919.7300	Europe +33.(0)1.3904.6880	Japan +81.(0)77.565.6999
--	---------------------------------	------------------------------	-----------------------------

Certificate of Analysis

Cat. Nos. 634437, 634438, 634455, 634456

SMART-Seq HT Kit

Product Documents

Documents for our products are available for download at takarabio.com/manuals

The following documents apply to this product:

- SMART-Seq HT Kit User Manual

Quality Control Data

A sample kit from each lot of the SMART-Seq HT Kit was functionally tested as follows: 10 pg of Control Total RNA was subjected to one-step cDNA synthesis and amplification as described in the SMART-Seq HT Kit User Manual. The first-strand cDNA was amplified using 17 cycles during the one-step RT-PCR reaction. PCR products were purified with a NucleoMag NGS Clean-up and Size Select kit (Takara Bio USA, Catalog No. 744970) and resuspended in 17 µl of Elution Buffer. 1 µl of the PCR reaction was analyzed with an Agilent 2100 Bioanalyzer and a High Sensitivity DNA Kit (Agilent, Cat. No. 5067-4626). The resulting analysis indicated a cDNA size range of approximately 400 bp to 10,000 bp. A peak was manually integrated from 300 bp to 9,000 bp, revealing a concentration of at least 200 pg/µl.

It is certified that this product meets the above specifications, as reviewed and approved by the Quality Department.

SMART-Seq® HT Kit

CATALOG NOS.

634437, 634438, 634455, 634456

NOTICE TO PURCHASER:

Our products are to be used for **Research Use Only**. They may not be used for any other purpose, including, but not limited to, use in humans, therapeutic or diagnostic use, or commercial use of any kind. Our products may not be transferred to third parties, resold, modified for resale, or used to manufacture commercial products or to provide a service to third parties without our prior written approval.

Your use of this product is also subject to compliance with the licensing requirements, listed below if applicable, and described on the product's web page at <http://www.takarabio.com>. It is your responsibility to review, understand and adhere to any restrictions imposed by these statements.

STATEMENT 275

SMART-Seq2 Technology. This product is sold under exclusive license from Ludwig Institute of Cancer Research, Ltd. and is covered by US Patent No. 10266894, Japanese Patent No. 6336080, and European Patent No. 3036336, and pending U.S. patent application and/or pending claims of foreign counterparts. For license information, please contact a Takara Bio USA, Inc. licensing representative by phone at 650.919.7320 or by e-mail at licensing@takarabio.com.

STATEMENT 388

This product is protected by Canadian Patent No. 2921603 and Japanese Patent Nos. 6336080 and 6603699 and corresponding US and foreign pending patents. For further license information, please contact a Takara Bio USA licensing representative by email at licensing@takarabio.com.

TRADEMARKS:

©2017 Takara Bio Inc. All Rights Reserved.

All trademarks are the property of Takara Bio Inc. or its affiliate(s) in the U.S. and/or other countries or their respective owners. Certain trademarks may not be registered in all jurisdictions.

Takara Bio USA, Inc.

1290 Terra Bella Avenue, Mountain View, CA 94043, USA

U.S. Technical Support: techUS@takarabio.com

United States/Canada

800.662.2566

Asia Pacific

+1.650.919.7300

Europe

+33.(0)1.3904.6880

Japan

+81.(0)77.565.6999

2/24/2021