

Certificate of analysis

Product Name/Description: TaKaRa PCR Human Papillomavirus Detection Set
Cat. #: 6602
Lot #: Specified on product label
Storage Conditions: -20 degrees C
Expiration Date: Specified on product label or package

Package Size: 50 Rxns

Package Contents:	1. HPVpF Primer	25 pmol/μl	100 μl
	2. HPVp16R Primer	25 pmol/μl	50 μl
	3. HPVp18R Primer	25 pmol/μl	50 μl
	4. HPVp33R Primer	25 pmol/μl	50 μl
	5. HPVb16 Probe	25 pmol/μl	10 μl
	6. HPVb18 Probe	25 pmol/μl	10 μl
	7. HPVb33 Probe	25 pmol/μl	10 μl
	8. Control Template HPV T16	1 ng/μl	50 μl
	9. Control Template HPV T18	1 ng/μl	50 μl
	10. Control Template HPV T33	1 ng/μl	50 μl

Quality Control Data:

PCR test:

- 1) Expected product (70 bp) was obtained by PCR using Control Template HPVT16 supplied in the set and the primer set (HPVpF and HPVp16R).
- 2) Expected product (70 bp) was obtained by PCR using Control Template HPVT18 supplied in the set and the primer set (HPVpF and HPVp18R).
- 3) Expected product (70 bp) was obtained by PCR using Control Template HPVT33 supplied in the set and the primer set (HPVpF and HPVp33R).
- 4) No specific amplification was confirmed by PCR using either of the primer sets and negative control (dH₂O).

It is certified that this product meets the above specifications, as reviewed and approved by the Quality Department of Takara Bio Inc.

Notice To Purchaser:

Please refer to the website.

Product Documents, Patent Markings, Trademarks, & Safety Information:

Refer to website for availability to download or contact your local office:
www.takara-bio.co.jp (Japanese)
www.takarabio.com (English)

Licence Statements:

Please visit our website and click the "license" button on the product's web page.

©2019 Takara Bio Inc. All Rights Reserved.

All trademarks are the property of Takara Bio Inc. or its affiliate(s) in the U.S. and/or other countries or their respective owners. Certain trademarks may not be registered in all jurisdictions.