

# Certificate of analysis

---

**Product Name/Description:** DNA-OFF®  
**Cat. #:** 9036  
**Lot #:** AOXP009  
**Storage Conditions:** Room temperature  
**Expiration Date:** February 2026

**Package Size:** 500 ml

**Package Contents:** 1. DNA-OFF® 500 ml

**Quality Control Data:**

Functional test:  
pUC19 template (3.4E + 08 copies) dried in PCR tubes is treated with DNA-OFF® or water control and then multiplex PCR is performed in the tubes using IL1-β template as an internal control on the PCR reaction alongside additional positive and negative controls. A successful result is indicated by the lack of a pUC19 amplicon in those tubes treated with DNA-OFF®.

Physical tests:  
The detergent blended into this product was confirmed by the bubbles formed when this product was shaken. The pH and the appearance of this product was 6-7 and clear, respectively.

It is certified that this product meets the above specifications, as reviewed and approved by the Quality Department of Takara Bio Inc.

**Notice To Purchaser:**

Please refer to the website.

**Product Documents, Patent Markings, Trademarks, & Safety Information:**

Refer to website for availability to download or contact your local office:

[www.takara-bio.co.jp](http://www.takara-bio.co.jp) (Japanese)

[www.takarabio.com](http://www.takarabio.com) (English)

**Licence Statements:**

Please visit our website and click the "license" button on the product's web page.

**©2019 Takara Bio Inc. All Rights Reserved.**

All trademarks are the property of Takara Bio Inc. or its affiliate(s) in the U.S. and/or other countries or their respective owners. Certain trademarks may not be registered in all jurisdictions.