

Certificate of Analysis

TAKARA

Product name/description : CultiLife™ 215 Culture bag

Cat. # : FU0005
Lot. # : 20200129
Storage Condition : Room temperature
Shipping Condition : Room temperature
Expiration Date : January 28, 2023
Package Size: 10 pieces

Quality Control Data :

| Test items | Specifications | Results |
|--|-------------------|-----------|
| 1) Appearance | Visual inspection | Passed |
| 2) Function test | | |
| 2-1) Leakage test | No leakage | Passed |
| 2-2) Fitting strength with tube | Not loose | Passed |
| 2-3) Transparency test | > 80% | 95% |
| 3) Counting Viable bacteria in the bag | < 25 cfu/bag | 0 cfu/bag |
| 4) Sterility Testing | Negative | Negative |
| 5) Endotoxin Testing | ≤ 0.25 EU/mL | Passed |

It is certified that this product meets above specifications.

Approved By :

Manager, Quality Assurance
TAKARA BIO INC.

Safety Information :

Please refer to our website for safety information :
<http://www.takara-bio.com>

Notice To Purchaser :

This product is not intended for human in vivo applications. It is the end user's responsibility to ensure that the final product meets the requirements of the application for which it is to be used.

This product may not be resold or transferred, modified for resale or transfer without written approval from TAKARA BIO INC.

If you require licenses for other use, please contact us by phone at +81 77 565 6973 or from our website.

<http://www.takara-bio.com>

Your use of this product is also subject to compliance with any applicable licensing requirements described on the product web page.

It is your responsibility to review, understand and adhere to any restrictions imposed by such statements.

All trademarks are the property of their respective owners. Certain trademarks may not be registered in all jurisdictions.

Country of origin: Japan