

Certificate of Analysis

TAKARA

Product name/description : GT-T610 (CultiLife™ Eva) Culture bag

Cat. # : FU0010
Lot. # : 20230718
Storage Condition : Room temperature
Shipping Condition : Room temperature
Expiration Date : July 17, 2026
Package Size: 10 pieces

Quality Control Data :

Test items	Specifications	Results
1) Appearance	Visual inspection	Passed
2) Function test		
2-1) Leakage test	No leakage	Passed
2-2) Fitting strength with tube	Not loose	Passed
2-3) Transparency test	> 80%	97%
3) Counting Viable bacteria in the bag	< 25 cfu/bag	0 cfu/bag
4) Sterility Testing	Negative	Negative
5) Endotoxin Testing	≤ 0.25 EU/mL	Passed

It is certified that this product meets above specifications.

Approved By :

Kinuko Nagamine

Manager, Quality Assurance
TAKARA BIO INC.

Safety Information :

Please refer to our website for safety information :
<http://www.takara-bio.com>

Notice To Purchaser :

This product is not intended for human in vivo applications. It is the end user's responsibility to ensure that the final product meets the requirements of the application for which it is to be used.

This product may not be resold or transferred, modified for resale or transfer without written approval from TAKARA BIO INC.

If you require licenses for other use, please contact us by phone at +81 77 565 6973 or from our website.

<http://www.takara-bio.com>

Your use of this product is also subject to compliance with any applicable licensing requirements described on the product web page.

It is your responsibility to review, understand and adhere to any restrictions imposed by such statements.

All trademarks are the property of their respective owners. Certain trademarks may not be registered in all jurisdictions.

Country of origin: Japan