

# Certificate of analysis

---

**Product Name/Description:** Takara FFPE DNA QC All-in-One Kit  
**Cat. #:** NN0001  
**Lot #:** Specified on product label  
**Storage Conditions:** -20 degrees C  
**Expiration Date:** Specified on product label or package

**Package Size:** 100 Rxns

<b>Package Contents:</b>	1. 2X Premix FFPE QC	2X	1 ml	x 2
	2. Short Assay Primer/Probe Mix	10X	200 µl	
	3. Long Assay Primer/Probe Mix	10X	200 µl	
	4. RNase Free H2O		1 ml	x 2
	5. DNA Standard FFPE QC	3.3x10 <sup>5</sup> copies/µl	100 µl	
	6. EASY Dilution (for Real Time PCR)		1 ml	x 2
	7. ROX Reference Dye II	50X	100 µl	

**Quality Control Data:**

Real-time PCR test:

The specific and quantitative amplification was confirmed by real-time PCR using a serial diluted solution ranging from  $6.6 \times 10^3$  to 26 copies of a standard DNA as a positive control template.

No specific amplification was detected by real-time PCR without the positive control.

It is certified that this product meets the above specifications, as reviewed and approved by the Quality Department of Takara Bio Inc.

**Notice To Purchaser:**

Please refer to the website.

**Product Documents, Patent Markings, Trademarks, & Safety Information:**

Refer to website for availability to download or contact your local office:

[www.takara-bio.co.jp](http://www.takara-bio.co.jp) (Japanese)

[www.takarabio.com](http://www.takarabio.com) (English)

**Licence Statements:**

Please visit our website and click the "license" button on the product's web page.

**©2023 Takara Bio Inc. All Rights Reserved.**

All trademarks are the property of Takara Bio Inc. or its affiliate(s) in the U.S. and/or other countries or their respective owners. Certain trademarks may not be registered in all jurisdictions.