



RUBICON GENOMICS

Certificate of Analysis

PicoPLEX™ WGA Kit	
Catalog No.	R30050
Lot No.	211704
Expiration Date	04/2017

Part Number	Component Name	Lot Number
BF00051	Cell Extraction Buffer	211682
BF00052	Extraction Enzyme Dilution Buffer	211683
EN00053	Cell Extraction Enzyme	211684
BF00054	Pre-Amp Buffer	211685
EN00055	Pre-Amp Enzyme	211686
BF00056	Amplification Buffer	211687
EN00057	Amplification Enzyme	211688
BF00058	Nuclease-Free Water	211689

This combination of components has been functionally validated.

Quality Control Testing		
Test Parameter	Test Specification	Result
Amplification Yield	Average amplified product yield of at least 2 micrograms from four 15 pg human genomic DNA samples.	Pass
Amplification Background	Real-Time Amplification curves from four no template samples are at least four cycles delayed relative to amplification curves from four 15 pg genomic samples.	Pass

Generated By: Nicole Palaski

Date: 02/01/2016

Rubicon Genomics, Inc., 4743 Venture Dr., Ann Arbor, MI 48108 USA
www.rubicongenomics.com

Tel: 734-677-4845