

DNA Unique Dual Index Kit - 24U Set A

Catalog No. Amount Lot Number

R400665 48 rxns Specified on product label.

Description

DNA unique dual index kits contain unique dual-indexed PCR primers designed for use with ThruPLEX® and PicoPLEX® kits to construct NGS libraries for multiplexed sequencing on Illumina® sequencers.

Package Contents

- 10 μl U1 DNA Unique Dual Index (5 μM)
- 10 μl U2 DNA Unique Dual Index (5 μM)
- 10 μl U3 DNA Unique Dual Index (5 μM)
- 10 μl U4 DNA Unique Dual Index (5 μM)
- 10 μl U5 DNA Unique Dual Index (5 μM)
- 10 μl U6 DNA Unique Dual Index (5 μM)
- 10 μl U7 DNA Unique Dual Index (5 μM)
- 10 μl U8 DNA Unique Dual Index (5 μM)
- 10 μl U9 DNA Unique Dual Index (5 μM)
- 10 μl U10 DNA Unique Dual Index (5 μM)
- 10 μl U11 DNA Unique Dual Index (5 μM)
- 10 μl U12 DNA Unique Dual Index (5 μM)
- 10 μl U13 DNA Unique Dual Index (5 μM)
- 10 μl U14 DNA Unique Dual Index (5 μM)
- 10 μl U15 DNA Unique Dual Index (5 μM)
- 10 μl U16 DNA Unique Dual Index (5 μM)
- 10 μl U17 DNA Unique Dual Index (5 μM)
- 10 μl U18 DNA Unique Dual Index (5 μM)
- 10 μl U19 DNA Unique Dual Index (5 μM)
- 10 μl U20 DNA Unique Dual Index (5 μM)
 10 μl U21 DNA Unique Dual Index (5 μM)
- 10 μl U22 DNA Unique Dual Index (5 μM)
- 10 μl U23 DNA Unique Dual Index (5 μM)
- 10 μl U24 DNA Unique Dual Index (5 μM)

Storage Conditions

• Store at -20° C.

Expiration Date

Specified on product label.

DNA Unique Dual Index Kit - 24U Set A

Shipping Conditions

Dry ice

Product Documents

Documents for our products are available for download at <u>takarabio.com/manuals</u> The following documents apply to this product:

• DNA Unique Dual Index Kit - 24U Protocol-At-A-Glance

Quality Control Data

All primers were PAGE-purified and verified by ESI analysis.

All primers were functionally tested with 200 pg of 200 bp sheared human genomic DNA using the ThruPLEX DNA-Seq Kit to produce indexed libraries. The libraries were pooled, purified, quantified, and then sequenced on the Illumina sequencing platform. The following parameters were measured for libraries prepared with indexed primers, which were required to meet the following specifications:

| Test Parameters | Result (%) |
|-----------------------------------|------------|
| qPCR end-point fluorescence (%CV) | ≤20 |
| % Coverage (%CV) | ≤20 |
| Read pairs examined (%CV) | ≤20 |
| Barcode balance as min/mean (%) | ≥40 |

It is certified that this product meets the above specifications, as reviewed and approved by the Quality Department.

(102523) Page 2 of 2



DNA Unique Dual Index Kit - 24U Set A

CATALOG NO.

R400665

NOTICE TO PURCHASER:

Our products are to be used for Research Use Only. They may not be used for any other purpose, including, but not limited to, use in humans, therapeutic or diagnostic use, or commercial use of any kind. Our products may not be transferred to third parties, resold, modified for resale, or used to manufacture commercial products or to provide a service to third parties without our prior written approval.

Your use of this product is also subject to compliance with the licensing requirements, listed below if applicable, and described on the product's web page at http://www.takarabio.com. It is your responsibility to review, understand and adhere to any restrictions imposed by these statements.

TRADEMARKS:

©2023 Takara Bio Inc. All Rights Reserved.

All trademarks are the property of Takara Bio Inc. or its affiliate(s) in the U.S. and/or other countries or their respective owners. Certain trademarks may not be registered in all jurisdictions.

Takara Bio USA, Inc.

2560 Orchard Parkway, San Jose, CA 95131, USA U.S. Technical Support: technical_support@takarabio.com

United States/Canada Asia Pacific **Europe** 800.662.2566 +33.(0)1.3904.6880 10/25/2023