

# DNA Unique Dual Index Kit - 24U Set D

Catalog No. R400668 Amount 48 rxns Lot Number Specified on product label.

# Description

DNA unique dual index kits contain unique dual-indexed PCR primers designed for use with ThruPLEX® and PicoPLEX® kits to construct NGS libraries for multiplexed sequencing on Illumina® sequencers.

### Package Contents

- 10  $\mu$ l U73 DNA Unique Dual Index (5  $\mu$ M)
- 10 μl U74 DNA Unique Dual Index (5 μM)
- 10  $\mu$ l U75 DNA Unique Dual Index (5  $\mu$ M)
- 10  $\mu$ l U76 DNA Unique Dual Index (5  $\mu$ M)
- 10 µl U77 DNA Unique Dual Index (5 µM)
- 10  $\mu$ l U78 DNA Unique Dual Index (5  $\mu$ M)
- 10 µl U79 DNA Unique Dual Index (5 µM)
- 10  $\mu$ l U80 DNA Unique Dual Index (5  $\mu$ M)
- 10 µl U81 DNA Unique Dual Index (5 µM)
- 10  $\mu$ l U82 DNA Unique Dual Index (5  $\mu$ M)
- 10 µl U83 DNA Unique Dual Index (5 µM)
- 10 µl U84 DNA Unique Dual Index (5 µM)
- 10 µl U85 DNA Unique Dual Index (5 µM)
- 10 µl U86 DNA Unique Dual Index (5 µM)
- 10 µl U87 DNA Unique Dual Index (5 µM)
- 10 µl U88 DNA Unique Dual Index (5 µM)
- 10 μl U89 DNA Unique Dual Index (5 μM)
- 10 μl U90 DNA Unique Dual Index (5 μM)
- 10 µl U91 DNA Unique Dual Index (5 µM)
- 10 μl U92 DNA Unique Dual Index (5 μM)
- 10 µl U93 DNA Unique Dual Index (5 µM)
- 10 μl U94 DNA Unique Dual Index (5 μM)
- 10  $\mu$ l U95 DNA Unique Dual Index (5  $\mu$ M)
- 10 µl U96 DNA Unique Dual Index (5 µM)

### **Storage Conditions**

• Store at  $-20^{\circ}$ C.

### **Expiration Date**

• Specified on product label.

# Certificate of Analysis

DNA Unique Dual Index Kit - 24U Set D

# **Shipping Conditions**

• Dry ice

# **Product Documents**

Documents for our products are available for download at <u>takarabio.com/manuals</u> The following documents apply to this product:

• DNA Unique Dual Index Kit - 24U Protocol-At-A-Glance

# **Quality Control Data**

All primers were PAGE-purified and verified by ESI analysis.

All primers were functionally tested with 200 pg of 200 bp sheared human genomic DNA using the ThruPLEX DNA-Seq Kit to produce indexed libraries. The libraries were pooled, purified, quantified, and then sequenced on the Illumina sequencing platform. The following parameters were measured for libraries prepared with indexed primers, which were required to meet the following specifications:

Test Parameters	Result (%)
qPCR end-point fluorescence (%CV)	≤20
% Coverage (%CV)	≤20
Read pairs examined (%CV)	≤20
Barcode balance as min/mean (%)	≥40

It is certified that this product meets the above specifications, as reviewed and approved by the Quality Department.



# DNA Unique Dual Index Kit - 24U Set D

### CATALOG NO.

### R400668

#### NOTICE TO PURCHASER:

Our products are to be used for **Research Use Only**. They may not be used for any other purpose, including, but not limited to, use in humans, therapeutic or diagnostic use, or commercial use of any kind. Our products may not be transferred to third parties, resold, modified for resale, or used to manufacture commercial products or to provide a service to third parties without our prior written approval.

Your use of this product is also subject to compliance with the licensing requirements, listed below if applicable, and described on the product's web page at <u>http://www.takarabio.com</u>. It is your responsibility to review, understand and adhere to any restrictions imposed by these statements.

#### **TRADEMARKS:**

#### ©2023 Takara Bio Inc. All Rights Reserved.

All trademarks are the property of Takara Bio Inc. or its affiliate(s) in the U.S. and/or other countries or their respective owners. Certain trademarks may not be registered in all jurisdictions.

Takara Bio USA, Inc.						
2560 Orchard Parkway, San Jose, CA 95131, USA						
U.S. Technical Support: technical_support@takarabio.com						
United States/Canada	Asia Pacific	Europe	Japan	10/25/2023		
800.662.2566	+1.650.919.7300	+33.(0)1.3904.6880	+81.(0)77.565.6999			