

I. Introduction

5X One Step PrimeScript™ III RT-qPCR Kit, no ROX, GPR enables the user to perform accurate one-step RT-qPCR on RNA samples. The highly concentrated master mix allows for flexibility in handling larger sample volumes. The entire protocol to detection takes less than one hour, enabling fast time to result.

II. Required Materials

This protocol applies to the following Takara Bio products:

- 5X One Step PrimeScript III RT-qPCR Kit, no ROX, GPR (Cat. Nos. 638337 & XU0002)

Additional materials required:

- Primers and probes
- Micropipette and tips (with hydrophobic filters)
- Vortex mixer
- Benchtop centrifuge for tubes or plates
- 1.5 ml Eppendorf tubes, 200 µl PCR tubes, or 200 µl PCR plates for sample preparation
- Tubes or plates for real-time PCR
- A real-time PCR machine (real-time systems that do not require ROX reference dye, such as CFX96 Real-Time PCR Detection System, Bio-Rad, etc.)

III. Protocol

1. On ice, prepare the RT-qPCR reaction mix for all reactions of the RNA sample, plus 10% of the total reaction mix volume for overage, using the table below.

Before adding the 5X One Step PrimeScript III RT-qPCR Mix, mix the contents of the tube by inverting several times and then spin down; the mix may appear cloudy after storage, but this does not affect its performance.

Use the reaction mix immediately.

NOTES:

- The RNA sample should not be added to the reaction mix; it will be added during Step 3.
- Refer to the “5X One Step PrimeScript III RT-qPCR Kit, no ROX, GPR User Manual” for more information on calculating the overage.

RT-qPCR reaction mix (per 25 µl reaction)

Reagent	Final concentration	Singleplex	Multiplex (N* targets)
5X One Step PrimeScript III RT-qPCR Mix, no ROX, GPR	1X	5 µl	5 µl
PCR Forward Primer (10 µM)	0.2 µM [†]	0.4 µl	N* x 0.4 µl
PCR Reverse Primer (10 µM)	0.2 µM [†]	0.4 µl	N* x 0.4 µl
Probe (10 µM)	0.2 µM [†]	0.4 µl	N* x 0.4 µl
RNA sample	—	2 µl [‡]	2 µl [‡]
RNase Free H ₂ O	—	Up to 25 µl [‡]	Up to 25 µl [‡]
Total volume per reaction		25 µl	25 µl

*Where N represents the total number of targets of interest. The total reaction volume will still be 25 µl, while 0.4 µl of each primer and probe will be added per target.

[†]For more information about primer, probe, and sample concentrations, please refer to the user manual.

[‡]Adjust the reaction volume according to the recommendations for the real-time PCR instrument used.

5X One Step PrimeScript™ III RT-qPCR Kit, no ROX, GPR Protocol-At-A-Glance

2. Add [25 µl – Volume of RNA Sample] of the reaction mixture into a PCR tube or a 96-well PCR plate.
3. Dispense the volume of RNA sample into the PCR tube or plate well.
4. Set up the thermal cycler and run the assay using the reaction conditions below:

52°C	5 min
95°C	10 sec
40 cycles:	
95°C	5 sec
60°C	30 sec

NOTE: Please follow the instruction manual of the real-time qPCR machine used. If the default setting does not work, perform manual analysis per the instruction manual.

5. After the reaction is complete, check the amplification curve. Confirm that the analytical parameter is correct and that the Ct value has been calculated.

Contact Us	
Customer Service/Ordering	Technical Support
tel: 800.662.2566 (toll-free)	tel: 800.662.2566 (toll-free)
fax: 800.424.1350 (toll-free)	fax: 800.424.1350 (toll-free)
web: takarabio.com/service	web: takarabio.com/support
e-mail: ordersUS@takarabio.com	e-mail: technical_support@takarabio.com

Notice to Purchaser

General Purpose Reagent

This product is a general purpose reagent intended **For Laboratory Use**. Outside of the United States, this product is intended for research use only unless otherwise stated. This product is not intended for a specific application or made for any particular clinical use. The performance characteristics of this product have not been fully established. It is the user's responsibility to validate the performance of the product, and any component thereof, for any particular use. Resale or transfer of this product, any component thereof, or any substance produced through use of this product, or any component thereof, to any third party is expressly forbidden. To obtain additional rights, please contact your local office or go to our website for additional information.

Your use of this product is also subject to compliance with the licensing requirements, listed below if applicable, and described on the product's web page at <http://www.takarabio.com>. It is your responsibility to review, understand and adhere to any restrictions imposed by these statements.

© 2022 Takara Bio Inc. All Rights Reserved.

All trademarks are the property of Takara Bio Inc. or its affiliate(s) in the U.S. and/or other countries or their respective owners. Certain trademarks may not be registered in all jurisdictions. Additional product, intellectual property, and restricted use information is available at takarabio.com.

This document has been reviewed and approved by the Quality Department.